

Assurity MRI™

Single Chamber Pacemaker



Merlin@home™
Transmitter
Compatible

Product Highlights

- MRI Ready device tested in combination with MR Conditional leads for full-body scans using a 1.5T and 3T (Tesla) field strength MRI Scanner.*
- Physician-preferred size and physiologic shape minimize pocket size.^{1,2}
- InvisiLink™ wireless telemetry system, in conjunction with the Merlin@home™ transmitter and Merlin.net™ Patient Care Network (PCN), allows for daily remote monitoring and follow-up.
- A suite of state-of-the-art features — such as automaticity, Ventricular AutoCapture™ pacing system for ventricular implants, ACap™ Confirm Feature, AF Suppression™ algorithm and AT/AF diagnostics for atrial implants, and SenseAbility™ sensing algorithm technology** — are designed to deliver optimal therapy for patients at implant and throughout their lives.

- Outstanding longevity provides 13.9 years of service life,³ which is supported by a 10-year warranty.⁴
- Six-month ERI-EOL interval.

*MRI Scan Parameters in MRI-Ready Systems manual

**Based upon the lead chamber selected

Ordering Information

MRI Pacing System

MODEL NUMBER	DESCRIPTION	DIMENSIONS (H × W × T, MM)	WEIGHT (G)	VOLUME (CC)	CONNECTOR
PM1272	Assurity MRI™ Pacemaker	47 × 50 × 6	20	10.4 (± 0.5)	IS-1

MODEL NUMBER	DESCRIPTION	INSULATION	FIXATION	MINIMUM INTRODUCER (F)	CONNECTOR	LENGTH (CM)
LPA1231	UltiPace™ Pacing Lead	Optim™	Ext/Ret helix	6	IS-1 bipolar	46, 52, 58, 65
2088TC	Tendril™ STS Pacing Lead	Optim™	Ext/Ret helix	6	IS-1 bipolar	46, 52, 58, 65***, 100***

***Not MR Conditional



Product Specifications

PHYSICAL SPECIFICATIONS	
Model	PM1272
Telemetry	RF
Dimensions (mm)	47 × 50 × 6
Weight (g)	20
Volume (cc)	10.4
Connector	IS-1
Remote Monitoring	Compatible with Merlin@home™ Transmitter

RATE / TIMING	
Parameter	Settings
Atrial or Ventricular Pace/ Sense Refractory (ms)	125; 160–400 in steps of 30; 440; 470°
Base Rate (bpm)	30–130 in steps of 5; 140–170 in steps of 10
Mode	VOO(R); VVI(R); VVT(R); Pacing Off, AOO(R); AAI(R); AAT(R)
Hysteresis Rate (bpm)	Off; 30 ⁷ –150 in steps of 5
<i>Search Interval (bpm)</i>	Off; 1; 5; 10; 15; 30
<i>Cycle Count</i>	1–16 by 1
<i>Intervention Rate (bpm)</i>	Off; 80–120 in steps of 10; Intrinsic +0; Intrinsic+10; Intrinsic +20; Intrinsic +30; Same as Base Rate
<i>Intervention Duration (min)</i>	1–10 in 1 minute intervals
<i>Recovery Time</i>	Fast; Medium; Slow; Very Slow
Rest Rate (bpm)	Off; 30–150 in steps of 5
Rate Responsive VREF	Off; Low; Medium; High
Shortest VREF	125–475 in steps of 25

RATE-MODULATED PARAMETERS	
Parameter	Settings
Sensor	On; Off; Passive
<i>Maximum Sensor Rate (bpm)</i>	80–150 in steps of 5; 160–180 in steps of 10
<i>Reaction Time</i>	Very Fast; Fast; Medium; Slow
<i>Recovery Time</i>	Fast; Medium; Slow; Very Slow
<i>Slope</i>	Auto (-1); Auto (+0); Auto (+1); Auto (+2); Auto (+3); 1–16 in steps of 1
<i>Threshold</i>	Auto (-0.5); Auto (+0.0); Auto (+0.5); Auto (+1.0); Auto (+1.5); Auto (+2.0); 1–7 in steps of 0.5

Product Specifications

OUTPUT/SENSING	
Parameter	Settings
ACap™ Confirm Feature ⁸	On; Off; Monitor
<i>Primary Pulse Configuration</i>	Bipolar
<i>Backup Pulse Configuration</i>	Bipolar
<i>Backup Pulse Amplitude (V)</i>	5.0
Search Interval (hours)	8; 24
A or V Pulse Amplitude (V)	0.25–4.0 in steps of 0.25; 4.5–7.5 in steps of 0.5
A or V Pulse Width (ms)	0.05; 0.1–1.5 in steps of 0.1
A or V Pulse Configuration	Unipolar (tip-case); Bipolar (tip-ring)
A or V Sense Configuration	Unipolar Tip (tip-case); Bipolar (tip-ring); Unipolar Ring (ring-case)
Atrial Sensitivity (mV)	0.1–0.4 ⁹ in steps of 0.1; 0.5; 0.75–2.0 in steps of 0.25; 2.5–4.0 in steps of 0.5; 5.0 ¹⁰
V Sensitivity (mV)	0.5–5.0 in steps of 0.5; 6–10 in steps of 1.0; 12.5 ¹⁰
Ventricular AutoCapture™ Pacing System	On; Off
<i>Primary Pulse Configuration</i>	Unipolar; Bipolar
<i>Backup Pulse Configuration</i>	Unipolar; Bipolar
<i>Backup Pulse Amplitude (V)</i>	5.0 ¹¹
<i>Search Interval (hours)</i>	8; 24
SenseAbility™ Sensing	Off; On
Algorithm Technology	(Automatic sensitivity control adjustment for atrial or ventricular events) 0.2–1.0 in steps of 0.1
<i>A Max Sensitivity (mV)</i>	0.2–2.0 in steps of 0.1
<i>V Max Sensitivity (mV)</i>	(Atrial and Ventricular Post-Sense) 50; 62.5; 75; 100%
<i>Threshold Start</i>	(Atrial Post-Pace) 0.2–3.0 in steps of 0.1 mV (Ventricular Post-Pace) Auto; 0.2–3.0 in steps of 0.1 mV (Atrial and Ventricular Post-Sense) 0; 30; 60; 95; 125; 160; 190; 220
<i>Decay Delay (ms)</i>	(Atrial Post-Pace) 0; 30; 60; 95; 125; 160; 190; 220 (Ventricular Post-Pace) Auto; 0; 30; 60; 95; 125; 160; 190; 220

MRI SETTINGS	
Parameter	Settings
MRI Mode	VOO or AOO (as applicable); Pacing Off
MRI Base Rate	85 bpm; 30–120 bpm in steps of 5 bpm
MRI Pulse Configuration	Bipolar
MRI Pulse Amplitude	5.0 V; 7.5 V
MRI Pulse Width	1.0 ms

MRI SCAN PARAMETERS™			
Lead Model	Magnet (Tesla)	RF Transmit Conditions	Scan Region
Tendril™ STS Pacing Lead	1.5 T / 3 T	Normal Operating Mode	Full-body
2088TC (lead lengths: 46, 52, 58 cm)			
UltiPace™ Pacing Lead	1.5 T / 3 T		
LPA1231 (Lead lengths 46, 52, 58, 65 cm)			

™™ For additional information about MR Conditional pacemakers and leads, including warnings, precautions, adverse conditions to MRI scanning and potential adverse events, please refer to the MRI-Ready Systems Manual at [medical.abbott/manuals](https://www.medical.abbott/manuals) or check our MRI Ready resources at [cardiovascular.abbott/mrireaddy](https://www.cardiovascular.abbott/mrireaddy).

Product Specifications

AF MANAGEMENT	
Parameter	Settings
AF Suppression™ Algorithm	Off; On (Atrial implants only)
Lower Rate Overdrive (bpm)	10 ¹¹
Upper Rate Overdrive (bpm)	5 ¹¹
No. of Overdrive Pacing Cycles	15–40 in steps of 5
Rate Recovery (ms)	8; 12 ¹¹
Maximum AF Suppression Rate (bpm)	80–150 in steps of 5; 160–180 in steps of 10
Atrial Tachycardia Detection Rate (bpm)	110–200 in steps of 10; 225–300 in steps of 25

STORED ELECTROGRAMS	
Parameter	Settings
Options	
<i>Priority Options</i>	Off; Low; High
<i>Channel</i>	1; 2; 3
Triggers	
<i>Magnet Response</i>	Off; Low; High
<i>High Ventricular Rate</i>	Off; Low; High
<i>Rate (bpm)</i>	125–300 in steps of 25
<i>No. of Consecutive Cycles</i>	2; 3; 4; 5; 10; 15; 20
<i>Advanced Hysteresis</i>	Off; Low; High
<i>Noise Reversion</i>	Off; Low; High
<i>High ventricular rate can alternately be high atrial rate; they use the same sub-parameters.</i>	

OTHER	
Parameter	Settings
Lead Monitoring	Monitor; Auto Polarity Switch
<i>V Low Impedance Limit (Ω)</i>	100–500 in steps of 25
<i>V High Impedance Limit (Ω)</i>	750–2500 in steps of 250; 3000
<i>Atrial limits apply when implanted in the atrium.</i>	
Lead Type	Uncoded; Unipolar; Bipolar
Magnet Response	Off; Battery Test
NIPS Options	
<i>Stimulation Chamber</i>	Atrial or Ventricular
<i>Coupling Interval (ms)</i>	100–800 in steps of 10
<i>S1 Count</i>	2–25 in steps of 1
<i>S1¹²; S2; S3 and S4 Cycle (ms)</i>	100–800 in steps of 10 (Fixed or Adaptive)
Diagnostic Trends	AT/AF Activity ⁸ ; Exercise; Lead Impedance; R (or P) Wave; V (or A) Threshold

References

1. Abbott. Data on file. Report 60048640. Market Research Report: Pacemaker Size and Shape.
2. Rajappan K. Permanent pacemaker implantation technique: Part I. Heart. 2009;95(3):259-264.

End Notes

3. A,V = 2.5 V @ 0.4 ms; 500 ohms; 100% VVI pacing @ 60 bpm; AutoCapture™ Pacing System OFF; SEGMs ON.
4. Terms and conditions apply; refer to the warranty for details.
5. ± 0.5 cc
6. Programming options dependent on pacing mode.
7. The highest available setting for hysteresis rate will be 5 bpm below the programmed base rate.
8. Atrial implants only.
9. Values 0.1–0.4 not available in unipolar sense configuration.
10. Sensitivity is with respect to a 20 ms haversine test signal.
11. This parameter is not programmable.
12. S1 burst cycle is applied at the preprogrammed S1 cycle length.

Rx Only

Brief Summary: Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use.

Indications: Implantation is indicated in one or more of the following permanent conditions: syncope, presyncope, fatigue, disorientation due to arrhythmia/bradycardia or any combination of those symptoms. **Rate-Modulated Pacing** is indicated for patients with chronotropic incompetence, and for those who would benefit from increased stimulation rates concurrent with physical activity. **Dual-Chamber Pacing** is indicated for those patients exhibiting: sick sinus syndrome, chronic, symptomatic second- and third-degree AV block, recurrent Adams-Stokes syndrome, symptomatic bilateral bundle branch block when tachyarrhythmia and other causes have been ruled out. **Atrial Pacing** is indicated for patients with sinus node dysfunction and normal AV and intraventricular conduction systems. **Ventricular Pacing** is indicated for patients with significant bradycardia and normal sinus rhythm with only rare episodes of A-V block or sinus arrest, chronic atrial fibrillation, severe physical disability. **AF Suppression™ algorithm** is indicated for suppression of paroxysmal or persistent atrial fibrillation episodes in patients with one or more of the above pacing indications.

Contraindications: **Dual-chamber pulse generators** are contraindicated in patients with an implanted cardioverter-defibrillator. **Rate-Adaptive Pacing** may be inappropriate for patients who experience angina or other symptoms of myocardial dysfunction at higher sensor-driven rates. An appropriate Maximum Sensor Rate should be selected based on assessment of the highest stimulation rate tolerated by the patient. **AF Suppression** stimulation is not recommended in patients who cannot tolerate high atrial-rate stimulation. **Dual-Chamber Pacing**, though not contraindicated for patients with chronic atrial flutter, chronic atrial fibrillation, or silent atria, may provide no benefit beyond that of single-chamber pacing in such patients. **Single-Chamber Ventricular Demand Pacing** is relatively contraindicated in patients who have demonstrated pacemaker syndrome, have retrograde VA conduction, or suffer a drop in arterial blood pressure with the onset of ventricular pacing. **Single-Chamber Atrial Pacing** is relatively contraindicated in patients who have demonstrated compromise of AV conduction.

Potential Adverse Events: The following are potential complications associated with the use of any pacing system: air embolism; body rejection phenomena; cardiac tamponade or perforation; hematoma, bleeding hematoma, seroma; formation of fibrotic tissue, local tissue reaction; inability to interrogate or program due to programmer or device malfunction; infection; erosion; interruption of desired pulse generator function due to electrical interference, either electromyogenic or electromagnetic; lead malfunction due to conductor fracture or insulation degradation; loss of capture or sensing due to lead dislodgement or reaction at the electrode/tissue interface; loss of desired pacing and/or sensing due to lead displacement, body reaction at electrode interface, or lead malfunction (fracture or damage to insulation); loss of normal device function due to battery failure or component malfunction; pacemaker migration or pocket erosion; pectoral muscle or diaphragmatic stimulation; phrenic nerve stimulation; pneumothorax/hemothorax; device migration and pocket erosion; endocarditis; excessive bleeding; induced atrial or ventricular arrhythmias; myocardial irritability; pericardial effusion; pericardial rub; pulmonary edema; rise in threshold and exit block; valve damage; death.

Refer to the User's Manual for detailed indications, contraindications, warnings, precautions and potential adverse events.

Abbott
15900 Valley View Court, Sylmar, CA 91342
Tel: +1 818 362 6822
Abbott.com

™ Indicates a trademark of the Abbott group of companies.

‡ Indicates a third party trademark, which is property of its respective owner.

© 2024 Abbott. All Rights Reserved.
MAT-2100149 v3.0 | Item approved for U.S. only.

